

GeneMAGNET PCR / DNA Clean-Up Purification Kit

Kit for purification of PCR products / DNA after enzymatic reactions

○ **Cat. no. E3420**

EURx Ltd. 80-297 Gdansk Poland
ul. Przyrodnikow 3, NIP 957-07-05-191
KRS 0000202039, www.eurx.com.pl
orders: email: orders@eurx.com.pl
tel. +48 58 524 06 97, fax +48 58 341 74 23





Table of Contents

Introductory Notes..... 3
 Equipment and reagents to be supplied by the user 3
Protocol..... 4
Safety Information 5

Content	96 preps E3420-01	Storage/Stability
Orange DX	48 ml	15-25°C
Wash DX1	60 ml	15-25°C
Wash DX2	80 ml	15-25°C
Elution	20 ml	15-25°C
Magnetic Beads	1000 µl	2-8°C
Protocol	1	

Introductory Notes

NOTE 1 • Kit Specification. The kit is suitable for fast cleanup of DNA fragments from PCR and other enzymatic reactions. This kit selectively removes primers below 40 nt and double-stranded DNA below 20 bp. However, common short by-products of not optimal or problematic PCR, known as primer-dimers, also consist of double-stranded DNA. They are produced from self-annealed and extended primers and co-migrate on a gel along with unincorporated single-stranded DNA primers. These double-stranded DNA artefacts co-purify with an expected PCR product, if their length exceeds 20 bp. If the removal of primer-dimers is necessary, we recommend PCR reaction optimization and/or agarose gel electrophoresis followed by isolation of PCR product using our spin-column GeneMATRIX Agarose-Out Purification Kit E3540.

NOTE 2 • Maximum Sample Amount.

Sample volume can be scaled up, the limit is the vial maximal volume. DNA cleanup can be performed in 1.5-2 ml Eppendorf tubes or on 96-well plates with working volume 800 µl.

NOTE 3 • Kit Compounds Storage. Once the kit is unpacked, store components at room temperature except Magnetic Beads that should be store in 2-8°C. In case of buffer Orange DX and Wash DX1 precipitation, simply warm up in 37°C water bath, until clarified.

NOTE 4 • Maintaining Good Working Practice. All solutions should be kept tightly closed to avoid evaporation and resulting concentration changes of buffer components. To obtain high quality DNA, stick carefully to the protocol provided below.

Equipment and reagents to be supplied by the user

1. Magnetic stand E0361 for 16 tubes, E0362 for 24 tubes, E0363 for 96-well plate. To be purchased separately.
2. Disposable gloves, pipettes, sterile pipette tips, sterile 1.5-2 ml tubes or 96-well plates (well volume at least 800 µl), a heating block capable of incubation at 56°C, vortex and laboratory rack for the tubes.



Protocol

1. Add 3 volumes of orange-coloured **Orange DX** buffer to 1 volume of the DNA sample and mix.
 - For example, add 300 μl of Orange DX buffer to 100 μl DNA sample.
2. Resuspend **Magnetic Beads** before removing them from the storage tube by vortexing. Add 10 μl of resuspended **Magnetic Beads** to the sample.
3. Mix by vortexing or pipeting for 1 min.
4. Separate the **Magnetic Beads** against the side of the tubes/wells. Wait until all the beads have been attached to the magnets.
5. Remove and discard the supernatant by pipeting.
6. Remove the magnetic stand/transfer tubes to the laboratory rack, add 500 μl of **Wash DX1** and mix by pipeting or vortexing for 10 s.
7. Separate the **Magnetic Beads** against the side of the tubes/wells. Wait until all the beads have been attached to the magnets. Remove and discard the supernatant by pipeting.
8. Remove the magnetic stand/transfer tubes to the laboratory rack, add 300 μl of **Wash DX2** and mix by pipeting or vortexing for 10 s.
9. Repeat steps 7-8.
10. Separate the **Magnetic Beads** against the side of the tubes/wells. Wait until all the beads have been attached to the magnets. Remove and discard the supernatant by pipeting.
 - Remove all the remaining Wash DX2 solution from the bottom of the tube/well.
11. Leave the open tubes/plate in magnetic stand and air dry the beads for 15 min.
 - Wash DX2 contains alcohol, make sure all the solution evaporates before proceeding to step 12.
12. Add 50-100 μl of **Elution** buffer to the tube/well and mix thoroughly by pipeting or vortexing and incubate 2-5 min.
13. Separate the **Magnetic Beads** against the side of the wells. After all the beads have been attached to the magnets transfer the supernatant containing the purified DNA to a suitable tube/plate. DNA is ready for analysis/manipulations. Isolated DNA can be stored either at 2-8°C or at -20°C.

Safety Information

Orange DX

Danger



H302+H332 Harmful if swallowed or if inhaled.

H315 Causes skin irritation.

H319 Causes serious eye irritation.

H334 May cause allergy or asthma symptoms or breathing difficulties if inhaled.

H317 May cause an allergic skin reaction.

P280 Wear protective gloves/protective clothing/eye protection/face protection.

P284 [In case of inadequate ventilation] wear respiratory protection.

P301+P312 If swallowed: call a poison center/ doctor if you feel unwell.

P304+P340 If inhaled: remove person to fresh air and keep comfortable for breathing.

P305+P351+P338 If in eyes: rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P333+P313 If skin irritation or rash occurs: get medical advice/attention.



Wash DX1

Warning



H226 Flammable liquid and vapour.

H302+H332 Harmful if swallowed or if inhaled.

H315 Causes skin irritation.

H319 Causes serious eye irritation.

P210 Keep away from heat, hot surfaces, sparks, open flames and other ignition sources. No smoking.

P280 Wear protective gloves/protective clothing/eye protection/face protection.

P301+P312 If swallowed: call a poison center/ doctor if you feel unwell.

P302+P352 If on skin: wash with plenty of water.

P304+P340 If inhaled: remove person to fresh air and keep comfortable for breathing.

P305+P351+P338 If in eyes: rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.



Wash DX2

Danger



H225 Highly flammable liquid and vapour.

H319 Causes serious eye irritation.

P210 Keep away from heat, hot surfaces, sparks, open flames and other ignition sources. No smoking.

P280 Wear protective gloves/protective clothing/eye protection/face protection.

P305+P351+P338 If in eyes: rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P403+P235 Store in a well-ventilated place. Keep cool.

P337+P313 If eye irritation persists: get medical advice/ attention.



**SELECTION OF THE KITS
DEPENDING ON THE TYPE
OF ISOLATED MATERIAL**

		ISOLATION OF DNA																					
		E3600	E3885	E3940	E3880	E3910	E3945	E3960	E3955	E3925	E3920	E3995	E3935	E3900	E3965	E3915	E3970	E3975	E3930	E3950	E3951		
		AVAILABLE NUMBER OF ISOLATION (PREPS)																					
		50 150	25 100	50 150	50 150	25 100	50 150	25 50	50 150	25 100	50 150	50 150	50 150	50 150	50 150	25 100	50 100	50 100	25 100	50 150	50 150		
DNA	GENOMIC	BACTERIA	●		●																	●	
		YEAST	●		●																		
		CELL CULTURE								●												●	●
		PLANT												●									
		FUNGI												●									
		PLANT RICH IN POLYSACCHARIDES ¹												●									
		BLOOD														●							
		SOIL																	●				
		STOOL																		●			
		SWAB																			●		
		ANIMAL TISSUES																				●	●
		FFPE TISSUE SECTIONS																				●	●
		RODENT TAILS																				●	●
		HAIR																				●	●
		INSECTS																				●	●
		URINE																				●	●
		BONE									●												
		BIOLOGICAL TRACES						●															
	FOOD										●												
	PLASMID	BACTERIA												●	●								
YEAST					●																		
ISOLATION FROM AGAROSE GELS				●			●																
PURIFICATION OF PCR PRODUCTS / DNA AFTER ENZYMATIC REACTIONS		●					●					●					●						

All kits contain buffers WASH in ready to use form

1. Additionally required lyse CT buffer (E0324)

2. Kit for creation of emulsions and subsequent DNA purification.

GeneMAGNET PCR / DNA Clean-Up Purification Kit is fast cleanup of DNA fragments from PCR and other enzymatic reactions.

Fragment of sizes from approximately 100 bp to over 15 kb can be obtained in ultrapure form. Contaminants such as: ethidium bromide, primers (below 40 nt), short double-stranded DNA (below 20 bp), RNA, Taq DNA Polymerase, Pfu DNA Polymerase, endo- and exonucleases, DNA-binding and modifying proteins, BSA and other enzymes/proteins, lipids, endotoxins, dyes, detergents, nucleotides, radio- and chemical

labels, EDTA, problematic restriction and ligation inhibitors, buffers and salts are effectively removed. Coloured binding buffer is very helpful in simultaneous processing of multiple samples. DNA is then eluted in low salt buffer, e.g.: Tris-HCl, TE or water. Isolated DNA is ready for downstream applications without the need for ethanol precipitation.

- **GeneMAGNET line is based on the use of silica paramagnetic beads (Magnetic Beads) for selective binding of RNA and DNA. The use of specially designed binding and washing buffers enables the efficient purification of highly pure nucleic acids.**



EURx Ltd. 80-297 Gdansk Poland
ul. Przyrodnikow 3, NIP 957-07-05-191
KRS 0000202039, www.eurx.com.pl
orders: email: orders@eurx.com.pl
tel. +48 58 524 06 97, fax +48 58 341 74 23

